Contents lists available at ScienceDirect

Pervasive and Mobile Computing

journal homepage: www.elsevier.com/locate/pmc

Health and sleep nursing assistant for real-time, contactless, and non-invasive monitoring *



Maria Valero^{a,*}, Jose Clemente^{b,*}, Fangyu Li^c, WenZhan Song^b

^a College of Computing and Software Engineering, Kennesaw State University, Marietta, USA

^b Center for Cyber-Physical Systems, University of Georgia, Athens, USA

^c Department of Electrical and Computer Engineering, Kennesaw State University, Marietta, USA

ARTICLE INFO

Article history: Received 20 July 2020 Received in revised form 5 May 2021 Accepted 29 May 2021 Available online 2 June 2021

Keywords: Sleep monitoring Vibration Fall detection Advanced signal processing Machine learning

ABSTRACT

In this paper, we introduce a novel health and sleep nursing assistant called "Helena" for real-time, contactless, and non-invasive monitoring that can be mounted on a bed frame to continuously monitor sleep activities (entry/exit of bed, movement, and posture changes), vital signs (heart rate and respiration rate), and falls from bed in a pervasive computing manner. The smart sensor senses bed vibrations generated by body movements to characterize sleep activities and vital signs based on advanced signal processing and machine learning methods. The device can provide information about sleep patterns, generate real-time results, and support continuous sleep assessment and health tracking. The novel method for detecting falls from bed has not been attempted before and represents a life-changing for high-risk communities, such as seniors. Comprehensive tests and validations were conducted to evaluate system performances using FDA approved and wearable devices. Our system has an accuracy of 99.5% detecting on-bed (entries), 99.73% detecting off-bed (exits), 97.92% detecting movements on the bed, 92.08% detecting posture changes, and 97% detecting falls from bed. The system estimation of heart rate (HR) ranged ± 2.41 beats-per-minute compared to Apple Watch Series 4, while the respiration rate (RR) ranged ± 0.89 respiration-per-minute compared to an FDA (Food and Drug Administration) oximeter and a metronome.

© 2021 Elsevier B.V. All rights reserved.

1. Introduction

Currently, 654 million people are 65 years and older worldwide [1], and around 47.8 million live in the United States [2]. Within those people, nearly 26% live alone at home and 18% in senior healthcare facilities or similar according to U.S Census Bureau [2]. The fact, that seniors are living longer [2] and increasingly want to stay in senior care facilities or their homes as they age, means more caregivers will be needed. In the U.S. alone, by 2030 the shortfall will be 151,000 paid direct-care workers and 3.8 million unpaid family caregivers. Smart sensors imbued with artificial intelligence (AI) would assist nursing healthcare workers to improve the quality of care while reducing the cost. However, for this population, there is a lack of contactless, non-invasive, and privacy-preserving solutions for monitoring sleep activities. Most of the

https://doi.org/10.1016/j.pmcj.2021.101422 1574-1192/© 2021 Elsevier B.V. All rights reserved.

This research is partially supported by NSF-STTR-1940864 and NSF-1663709.
 Corresponding authors.

E-mail addresses: mvalero2@kennesaw.edu (M. Valero), jclementes@uga.edu (J. Clemente), fli6@kennesaw.edu (F. Li), wsong@uga.edu (W. Song).

URLs: http://iotas.kennesaw.edu/ (M. Valero), https://sites.google.com/view/fangyuli (F. Li), https://sensorweb.engr.uga.edu/index.php/song/ (W. Song).

monitoring solutions in the market are conceived to measure just vital signs (heart rate) with a wearable apparatus/gadget, and those do not detect other important activities like lack of movement on the bed and falls from bed. Other devices require the person's actions, like pressing a button when a fall happens, but those cannot be activated if the person loses consciousness. On the other hand, the use of cameras to monitor sleep generally raises privacy concerns.

When senior people require a sleep assessment, the standard technology is polysomnography (PSG) [3], which records multiple physiological signals and is generally performed in a sleep clinic. However, unless medically needed, this kind of assessment is costly for families. Most commercial solutions include wrist devices that have been shown to estimate nocturnal sleep duration and sleep-wake patterns [4]. However, in the elderly population, the major concern is that they will forget to use the device. Moreover, some of these wearable sleep monitoring products require user intervention for indicating initial and final times of measurements (such as in smart watches like Samsung), or they only estimate partial vital signs (such as FitBit, which does not consider respiration or blood pressure). Even more sophisticated wearable devices like Philips Alice Pdx and Sleep Pace require the use of uncomfortable devices on the chest and eyes, which is inconvenient for long term analysis [5,6]. Then a contactless and non-intrusive solution could be ideal for sleep monitoring and assessment. In the market, multiple contactless devices have been used for sleep monitoring, and they are usually placed under mattresses [7–9]. However, they do not detect changes in body posture on the bed (like Beddit Sleep Monitoring), or calculate important vital signs such as blood pressure (like Early Sense), and are too expensive for family use [8].

This paper introduces a novel health and sleep nursing assistant called "Helena" for real-time, contactless, and noninvasive monitoring that can be mounted on a bed frame to continuously monitor sleep activities (entry/exit of bed, movement, and posture changes) and vital signs (heart rate, respiration rate), and falls from bed. The technology includes a small yet smart vibration sensor that can be easily installed and a user-friendly graphic interface that can be paired with smart mobile devices to monitor results through a software application. Vibration sensors have been widely used in smart home applications [10–13]. The smart sensor senses bed vibrations generated by body movements and sleep activities and then infers sleep activities and vital signs based on advanced signal processing and machine learning methods. The system provides information about sleep patterns, generates real-time results, and supports continuous sleep assessment and health tracking.

The main contributions of our work are:

- A real-time, contactless, and non-invasive sleep monitoring system based on bed vibration, which allows estimation of heart and respiration rates as well as sleep activities like on/off bed detection, movements, posture changes, and for the first time falls from bed. The system allows sending alerts to smart devices when sudden changes occur, for example, the person has too many or too few movements, a fall down happens, or the person entries/leaves the bed. After the first 30 s of data (to construct the system's sliding window), the proposed system can estimate heart rate and respiration rate every 3 s, and detect movements, posture changes, and fall downs every 5 s, which represents a reliable way to notify caregivers about events while care recipient is sleeping.
- Accurate, light and fast methods for heart and respiration rates, suitable for pervasive devices, that can be executed in real-time on a single-board computer.
- First work to propose a method to detect and alert falls from bed using only vibration signals.
- First work to propose a method to detect movement and postures changes based only on the vibration signal generated by the body on the bed.

In addition to the contributions, in this paper, we demonstrate the feasibility of the proposed system compared to other similar products in the market (Section 2). We perform controlled and uncontrolled test with multiple people, types of beds, types of houses, types of floors in the houses, etc; and we also demonstrate the usability of the system in a senior healthcare facility (Section 6). We also discuss the drawbacks of the system and potential improvements for future work (Section 7).

2. Related works

Sleep monitoring is extremely important, even a lifesaver, for people with undiscovered illness, which causes respiration and heart failures [14,15]. The respiration status can be monitored by breathing apparatuses [16], while the heart rate is typically measured by wearable devices [14]. However, those devices need body contact and are intrusive. Many people feel not comfortable wearing or forget to wear before sleep. Comparing with FDA approved wearable devices, our system can identify with high accuracy sleep activities and vital signs. Furthermore, to the best of our knowledge, this is the first work that proposes a method to detect falls from bed using only vibration signals. There are multiple contact-based devices for sleep monitoring. For example, wearable devices like *Apple Watch* [17], *Jawbone* 3 [18], *Basis Peak* [19], *Fitbit Surge* [20], and a number of other "smartwatch" devices that have been good in demonstrated their capabilities to measure heart rate; however, they lack deep study of activities and patterns behaviors of the person during night and require user intervention for correct performance.

Typically, contactless sleep sensing approaches are based on the assessment of body behaviors during sleep by instrumenting sensors under the mattress [21]. Most of them are based on force sensors [22], load cells [23], multi-channel infrared sensor-arrays [24], and pressure sensors [25]. Recently, new applications for smart environments are

explored, such as ambient floor vibration for indoor person localization [13,26], bed vibration for heart beating and breathing rate monitoring [22,27,28], etc. Besides, there has been rare work on sleep posture identification using body vibration patterns during sleep. Multiple studies have been conducted in the past five years to extract heart rate and respiration rate using vibration sensors. For example, Jia et al. [28] were the first to propose a bed-mounted geophonebased heartbeat monitoring. In this study, authors use an SM-24 Geophone Element [29], an amplifier [30], and an Analog-to-Digital Converter (ADC) placed under the mattress to extract the heartbeat of the person on the bed. In this work, the authors used Sample Auto-correlation Function (ACF) [31] to extract periodicity from the time series and a peak finding algorithm developed by Thomas C. O'Haver [32] to locate peaks in the sample ACF results. Those peaks were considered the heartbeats. Even though the work has proved the concept and the feasibility of the approach, it has been implemented in an ideal noise-free scenario for only research purposes, and it does not approach real-world interference like washer machines, air conditioners, etc. In 2018, Li et al. [15] proposed a methodology to extract HR and RR using a seismometer. For HR, the authors proposed a Local Maxima Statistics (LMS) method, and for RR, they proposed an Instantaneous Property from the Oscillatory Analysis. Visual comparisons between results and peaks seem to match, but no further error measurements were provided. Recently, multiple efforts have been made to apply machine learning and deep learning techniques to vibration signals to estimate vital signs. In 2020, Park et al. [33] presented a geophone-based sensing system for extracting electrocardiogram (ECG) patterns using heartbeat vibration through a bed mattress. The signal is first filtered by passing the signal through a bandpass filter to keep frequencies between 5-30 Hz. Then, the data passed on to a Bidirectional Long-Short Term Memory (Bi-LSTM) [34] deep learning model for ECG waveform estimation. The deep learning model architecture in [33] is designed using two-stacked Bi-LSTMs and three fully connected layers. Even though the results are promising, the approach is computational heavy to be applied in-situ inside the sensor as the proposed approach. Other works have focused on estimating breathing states. Choudhary et al. [35] presented a method to assess the human respiratory system by identifying degree-of-breathings, such as breathlessness, normal breathing, and long breathing. The input data are SCG signals. The authors extracted 15 statistically significant morphological features from the SCG cycle including heart-rate, beat energy, beat entropy, kurtosis [36], autocorrelation feature [37], IM/IC amplitude [36], maximum spectral amplitude, and beat spectral centroid. Then, a Stacked Autoencoder (SAE) deep learning architecture was employed for the identification of different respiratory-effort levels. The approach is also promising but require post-processing of data and could be hard to implement in real-time scenarios

On the commercial side, under-mattress devices like *Beddit* [38] and *Nokia Withings Sleep Tracking Pad* [7] measure sleep time, heart and respiration rates, sleep cycles; however, these measurements are based on a whole night average that is shown the next day, which means there is not real-time information. EarlySense device [39] is the most near to real-time sleep monitoring and has been developed to provide continuous monitoring of heart rate, respiration rate, and bed motion for patients in medical/surgical scenarios using a pressure sensor; however, the device is not too robust in posture changes and fall from the bed, and it is very expensive.

We conduct an extensive product analysis has been carried out in order to highlight Helena's innovations. Table 1 shows the different features available in Helena and each wearable and contactless opponent product in the market. Compared with other wearable devices, Helena is the only one that allows real-time recognition of sleep activities like bed entry/exit, movements, and posture change. Table 1 shows: (1) contactless devices (superscript ¹) vs. wearable devices (superscript ²); (2) price range (\$ means < \$150; \$\$ means between \$150 and \$500; \$\$\$ means between \$500 and \$1000; \$\$\$\$ means \$1000; (3) the different features: heart rate (HR), respiration rate (RR), entry/exit bed detection (Entr/Exit), movement detection (Mov), posture change detection (Post), and fall down detection (Fall). Note also that features marked x• are referred to estimations in real time. Also, the symbols \star and \circ in the Camera device means "non-privacy" and "features under development" respectively.

To the best of our knowledge, Helena is the first vibration-based health and sleep nursing monitoring system that estimates heart rate in respiration rate in real-time, and infers sleep activities like entry/exit of the bed, movements on the bed, and changes of the body postures on the bed. Most importantly, Helena is the only solution that also detects fall downs using vibration signals. This information allows sophisticated analysis of health and sleep, which can also be reported according to the user's preferences and alert configuration.

3. Overview

We use a smart sensor composed by a geophone (seismic sensor) [51,52], a digitizer board and a single-board computer (Raspberry Pi 3B) [53]. It is shown in Fig. 1. The geophone detects the velocity of the movements. The digitizer transforms analog vibration signals to digital that is read by the single-board computer. The single-board computer reads and processes the signal to estimate the vital signs and on bed activities. After multiple tests, we decide that the sensor must be installed in a range no larger than 40 centimeters from the heart location when the person lying down. The estimated results are sent to the Cloud used the single-board computer WiFi. Also, it is in charge of sensing alert messages via email or/and SMS and communicating with the smart assistant (Google Home or Amazon Echo).

Fig. 2 illustrates Helena architecture. The raw data sensed by the geophone is used to determine the bed status (on/off/sitting). If the person is detected on the bed, Helena cleans the noise and estimates HR and RR if there are no movements in the last 15 s. It is the time window used to calculate the vital signs with high accuracy. The vital signs are estimated every 3 s. When a movement is detected, we evaluate three scenarios (posture change, off bed, and fall from

Comparison	between	Helena	and	wearable	and	contactless	sleep	monitoring	devices.

	Price	rice Features							
		HR	RR	Entry/Exit	Movement	Posture	Fall		
Helena ¹	n/a	x•	x•	x•	x•	x•	x•		
Zeo ² [40]	\$\$\$	x•							
Apple Watch ² [17]	\$\$	x•	х						
Jawbone ² [18]	\$	х	х						
Basis Peak ² [19]	\$	х							
Microsoft band ² [41]	\$	х							
FitBit ² [20]	\$\$	x•							
Philips Alice ² [42]	\$\$\$\$	х	x•						
WatchPAT ² [43]	\$\$\$\$	x•	x•			х			
Cameras ^{1*°} [44,45]	\$\$		х	х	х	х			
Sleep Cycle ¹ [46]	\$	х			х				
Emfit ¹ [9]	\$\$	х	х	х	х				
Beddit ¹ [38]	\$\$	х	х	х					
Nokia Pad ¹ [7]	\$	х		х					
ResMed ¹ [47]	\$		х	х					
Smart Bedding ¹ [48]	n/a	х	х		х				
Sleepace Pad ¹ [49]	n/a	x•	x•		х				
EarlySense Live ¹ [50]	\$\$	x•	x•	х	х				
FarlySense ¹ [8]	2222	v•	v•	v	v				



Fig. 1. Helena unit prototype. The geophone senses the bed vibration signal, and the digitizer receives the analog information and creates a digital representation which is processed by the single board computer.

bed). We use 5 s of data before and after the event to calculate posture change. For fall from the bed, we analyze if after the movement there is an off-bed status. Then, the system performs a machine learning model to detect if the event is a fall from the bed. Otherwise, an off-bed event is set.

Inside the Raspberry PI, we have configured automatic services to collect, read the data from the geophone, and save the data inside the database (InfluxDB [54]). Another service that is running constantly, extracts the data from the database and sends it to the algorithm section where the different algorithms are executed. We use sliding windows to process the data on each algorithm. The reading of the data is done in real-time. The only delay is at the beginning of the readings, where the algorithms are not providing heart rate, respiration rate, or any other information until they collect the first 30 s of data. After that, the algorithms utilize this sliding window second by second to calculate the values and show the results. Note that there is no real delay, the system just does show results during the first 30 s of data collection. After that, the system updates second by second by utilizing the old 29 s with the new 1 s from the sensor.

Additionally, we develop a real-time visualization tool based on InfluxDB [54] and Grafana [55]. InfluxDB is an opensource time-series database for storing values associated with a time stamp. The raw data is stored in a local database, and the results are sent to the Cloud. Grafana is used to create different dashboards such as raw data, bed status (on/off/sitting), HR, RR, movement, posture change, and fall from the bed. Also, the system sends alerts and notifications configurable by the user (email, SMS, or through the smart assistant). Alerts are generated when a person leaves the bed or if a fall is identified.

4. Challenges

There are three main challenges to estimate sleep activities and vital signs using a vibration signal captured by a sensor attached on the bed frame:

• Ambient noises: In vibration-based systems, the main issue is the unpredictable noise generated by natural and human activities. Constant noise, like the one produced by an air-conditioner, washer/dryer, vacuum may affect the



Fig. 2. Helena architecture.

data quality and vital signs estimation. It is a challenge to detect and eliminate the noises produced by the different machines while preserving the signal produced by the heartbeat. Because Helena is designed for monitoring a single person, additional people or pets in bed may general noise.

- Event recognition: Vibration-based systems are highly sensitive and can detect other activities besides the heartbeat. A major challenge is to distinguish and classify the relevant events for estimating sleep activities, vital signs, and fall downs, and discard other non-related events.
- **Heartbeat estimation:** The heartbeat signal received by the sensor varies depending on multiple factors like height and type of mattress, sensor location, types of beds. Because heartbeat recognition is the base for many of our algorithms, recognizing heartbeats in different environments/people/setups is a major challenge of the proposed development.

5. Algorithm and system design

5.1. Signal enhancing

Since our goal is to realize daily sleep monitoring, the proposed system should be robust to different environments. Besides the random noise suppression, an adaptive machine vibration elimination is also required in the home environment to handle the noises from home appliances, such as AC (air conditioner), washer/dryer, vacuum, and so on. Based on the geophone with a 100 Hz sampling rate, the available frequency range is $0 \sim 50$ Hz, according to the Nyquist sampling theorem. A spectrum scanning method is first applied to remove the machine noises in the environment. We apply notch filters to suppress the noise components with iso-dominant-frequencies. For HR/RR estimation, because the typical HR is between 40 bpm (beat per minute) and 150 bpm and RR is between 12 rpm (respiration per minute) and 25 rpm, which means the target HR/RR should include 0.2 Hz to 2.5 Hz, only a small frequency range is needed. Thus, we apply a bandpass filter with 0.1 Hz low-cut frequency and 8 Hz high-cut frequency to extract target vibration signals. After the process, the machine noise is removed, and the signal retains the vibration signatures.

5.2. On/off-bed – bed exits and entrances

Knowing if the person is effectively lying down on the bed is crucial for other system estimation modules. We propose a multiple feature fusion(MFF) method, that consists of different metrics to estimate if the person has entered or exited the bed. We use a fusion of Spectral Entropy (G), Kurtosis (U) and Teager Energy Operator (T) to estimate the real energy when the person is on the bed. The SE treats the signal's normalized power distribution in the frequency domain as a probability distribution and calculates the Shannon entropy of it. The Shannon entropy in this context is the spectral entropy of the



Fig. 3. Data-fusion energy based entry/exit bed detection (bottom), which is obtained by MFF. Note that the unit data generated by the geophone (top) is voltage.

signal. For a signal x(n), the power spectrum is $S(m) = |X(m)|^2$, where X(m) is the discrete Fourier transform of x(n). The probability distribution P(m) is then $P(m) = \frac{S(m)}{\sum S(i)}$. Then, the spectral entropy is estimated as

$$G = -\sum_{m=1}^{N} P(m) \log_2 P(m).$$
(1)

Kurtosis is the fourth standardized moment, defined as:

$$U[X] = \mathbb{E}\left[\left(\frac{X-\mu}{\sigma}\right)^4\right] = \frac{\mu_4}{\sigma^4} = \frac{\mathbb{E}[(X-\mu)^4]}{(\mathbb{E}[(X-\mu)^2])^2},$$
(2)

where μ is the fourth central moment and σ is the standard deviation. Finally, The T can be driven from a second-order differential equation [56]. The total energy of oscillation (i.e., the sum of kinetic and potential energies) can be obtained from the following equation

$$T = \frac{1}{2}kx^2 + \frac{1}{2}mx^2,$$
(3)

where m is the mass of the oscillating body and k is the spring constant. Results can be seen in Fig. 3. It is possible to see the decreasing of the total fused energy when the person is on the bed.

The data-fusion energy (DFE) shown in Fig. 3 based on MFF can be obtained in the following way. Assume the number of measurement in an analysis window is I, the value DFE_i of the *i* measurement is

$$MFF_i = G_i \alpha + U_i \beta + T_i \gamma, \qquad i = 1..I,$$
(4)

where α , β , and γ are the corresponding coefficients of G, U and T, respectively. Therefore, the DFE_{*I*} of the analysis window *I* can be obtained using the following expression:

$$DFE_{I} \stackrel{\Delta}{=} \begin{bmatrix} DFE_{1} \\ DFE_{2} \\ . \\ . \\ DFE_{I} \end{bmatrix} = \begin{bmatrix} G_{1} \\ G_{2} \\ . \\ . \\ G_{I} \end{bmatrix} \alpha + \begin{bmatrix} U_{1} \\ U_{2} \\ . \\ . \\ U_{I} \end{bmatrix} \beta + \begin{bmatrix} T_{1} \\ T_{2} \\ . \\ . \\ T_{I} \end{bmatrix} \gamma.$$
(5)

Note that the DFE is obtained not based on a single point but a relatively long time interval to assure a smooth result. The DFE is defined between 0 and 100 to keep the simplicity. DFE can be viewed as the combination of "real" energy measurements of the person on bed. When DFE is below a established threshold, we can consider that the person is on bed. For our algorithm, after multiple experiments, we selected the value of the threshold is 12.

5.3. Movements and posture change

To detect movements, we first use the MFF method to determine if the person has entered the bed. After that, we observe that the body movement generates a strong signal (Fig. 6 in our previous work [15] while the respiration and



Fig. 4. Five seconds signal from four different postures.

heartbeat show an amplitude about 10⁵. Thus, based on the dramatic energy change, we can recognize the body movement using a local thresholding method:

$$Tr_m = \begin{cases} 1 & \text{if } s(t) \ge \lambda \max(s(z)), z \in (t - \tau, t) \\ 0 & \text{otherwise}, \end{cases}$$
(6)

where, s(t) is the signal in time t, λ is a threshold coefficient and τ is the time lag.

Based on the on bed movement detection results, we develop a posture change detection method because not all body movements mean sleep posture changes. For posture change detection, we analyze 5 s signal before and after the movement to compare the similarity between them. We calculate the similarity average of four metrics – the Spectral Entropy, Kurtosis, TEO, and Power Spectral Density (PSD) – characterized by

$$sim = sum_{i=0}^{-4} (1 - abs(A_i - B_i)/((A_i - B_i)/2))$$
⁽⁷⁾

where, A_i is the metric *i* value for the signal after the movement, and B_i corresponds the metric *i* value for the signal before the movement. After several tests, a threshold of *sim* < 0.85 is established to identify if there is a posture change. Fig. 4 shows the signals of different postures for one person.

5.4. HR/RR estimation

Estimating hearth rate (HR), measured in BPM_h directly from the data spectrum [28] is not accurate because the heartbeat waveform is not strictly periodical in reality as it can be seen in Fig. 5.

The local maxima search method may fail (i.e., generate unreasonable results) when there are noises and other interference. To improve HR estimation stability, we develop an envelope based HR estimation method. The envelope is a curve such that at each point it touches tangentially the signal. Fig. 5 shows an example of the envelope obtained to estimate in HR. The parametric equations of the envelope are given implicitly as U(x, u, C) = 0 and $U'_C(x, u, C) = 0$. Once the envelope is obtained, the peaks can be used to estimate the HR. To solve instabilities, a novel empirical truncated statistics analysis method is proposed to estimate HR. When local maxima of envelope are obtained, there are falsely picked peaks and some missing peaks. Those falsely picked peaks result in smaller period estimation, whereas the missed peaks lead to larger estimation results. Here, X is the interval between two sequential picked peaks. The heartbeat period within $(t - \frac{l_h}{2}, t + \frac{l_h}{2})$ is estimated as a truncated average:

$$E(X|F^{-1}(a) < X \leq F^{-1}(b)) = \frac{\int_{a}^{b} xg(x)dx}{F(b) - F(a)},$$
(8)
where, $g(x) = f(x)$ for $F^{-1}(a) < x \leq F^{-1}(b);$
 $g(x) = 0$, everywhere else;
 $F^{-1}(p) = \inf\{x : F(x) \ge p\}.$

The lower and upper bounds (a and b) are determined based on the local maxima detection performance. In our applications, 0.1 and 0.9 are chosen for a and b, respectively.

Commodity seismometer is insensitive to lower frequency measurements (usually lower than 0.3 Hz) [27,28], thus the respiratory rate BPM_r cannot be directly observed from seismic data. Previously, an amplitude-modulation approach is proposed to use the envelope to estimate carrier frequency [28]. However, the amplitude modulation of the recorded seismic signal is not stable. For that reason, we propose a novel technique to obtain the envelope. We use the HR peaks detected to generate the respiratory modulation signal by extrapolating them. Then we obtain the envelope of the signal and count the peaks to estimate the respiratory rate. Fig. 5 shows the envelope generated by the proposed method.

5.5. Fall from bed

To identify "fall from bed", we propose to use a machine learning-based approach. Based on a supervised machine learning algorithm, Support Vector Machine (SVM) [57], data are transformed using a kernel to find the optimal boundary



Fig. 5. Heart signal. First and second local maxima estimation from envelope to estimate HR and RR.



Fig. 6. Bed vibration signals for different events.

between the different classes. SVM and its derivatives have been widely used in biomedical engineering [13,58] because it is efficient and easy to implement on embedded devices. Before applying the machine learning method, we identify whether the detected event is a candidate to be a fall from the bed. In the stage of "falling from bed", the person should be laying or sitting on-bed, and then an event needs to be detected. When an off-bed is identified after the event, it becomes a candidate to be a fall from the bed. At that moment, the features are extracted and classified by the SVM to identify if the event is a fall from bed or another movement. All events used during training and tests are shown in Fig. 6.

5.5.1. Feature extraction

We compute features in the time and frequency domain. The signal events are normalized to get similar features for each type of event. The normalization is done by dividing the event by the energy average of its signal, where the energy is defined as

$$\mathcal{E} = \sum_{n=0}^{N} |s(n)|^2.$$
(9)

For fall down detection, we initially utilize multiple features in the time and frequency domain like for example entropy and location of maximum peaks for time domain and centroid and number of peaks in the frequency domain. After careful evaluation of the parameters in a data-driven process, we realize that the most contributing features are related to signal amplitude, event duration, amount of sub-events, and the last 20 peak information in the time domain due to nature of the fall signal. Fig. 7(left) shows the time domain features. In the frequency domain, we select Power Spectral Density (PSD) of the vibration signals, which exposes the power distribution for different frequencies. If s_i represents the signals, then the PSD can be defined as

$$PSD_i = 10\log_{10}\frac{abs(FFT(s_i))^2}{f_s * n}$$
(10)

where f_s is the sampling frequency, n is the number of samples of received signal s_i and FFT(·) is the Fast Fourier Transform operation. Fig. 7(right) shows the PSD feature used as a part of the input.

5.5.2. Selection

For experimental purposes, we generate 480 simulated falls from the bed. To have plenty of variety of signals, the simulated falls are performed by 3 people in 4 different bed falling 20 times for each bed-side. To choose the correct classifier that obtains high levels of accuracy and to have the ability to adapt to the limited computing resources of the computer board, we evaluate the accuracy and run-time of three different classifiers using the 480 falls from the bed. The tested classifiers are Gaussian Process (GP), K-Nearest Neighbors (KNN), and a one-class SVM. The data is split into 80 percent for training and 20 percent for testing. To test the classifiers, we also use other non-fall events like people changing position on the bed and people standing up from the bed. In order to get this null-class data, we generate 150



Fig. 7. Time and frequency domain features for fall from bed classification.



Fig. 8. Receiver Operating Characteristic (ROC) of the cross-validation procedure using 10 folds.

events of changing positions in bed and 150 events of standing up from bed from 5 different people. The process was conducted on 4 different Raspberry Pis for each classifier. The results are shown in Table 2.

All classifiers have an acceptable level of accuracy. Even though KNN has the highest accuracy and recall, there is no big difference between SVM and GP. On the other hand, all training times are acceptable because they are executed just one time. However, SVM outperforms the classification run-time. The run-time increases 3.2X times when GP classifier is used, and 25.45X times using KNN. Then, we selected SVM as a classification as the classifier model.

5.5.3. Validation

The k-fold cross-validation procedure is applied to evaluate the classification model. We use k-fold to estimate how the model is expected to perform when it makes predictions on data not used during the training of the model. The procedure splits the data into k groups, and for each group, it (i) takes one group as a test data set, (ii) takes the remaining groups as a training data set, (iii) fits and evaluates the model, and (iv) retains the evaluation score. We select 10-folds to cross-validate the model. We use AUC (Area Under the Curve) ROC (Receiver Operating Characteristics) to check and visualize the performance of our classification problem. ROC is a probability curve, and AUC represents the degree or measure of separability. They tell how much the model is capable of distinguishing between classes. An excellent model has an AUC near 100, which means it has a good measure of separability. Fig. 8 shows the AUC for all folds. The AUC average is 89.92, which is an acceptable measurement.

6. Experiments and evaluations

We performed three types of experiments to validate the bed system. The first one was done in a lab environment to set up parameters, find issues, and validate the system with different people. Then, we carried out a test in different environments outside the lab to test different beds, noise levels, and to validate the HR results comparing with an FDA approved device. Finally, we set up two units in a senior assisted living facility in Loganville, GA to test the system with senior people, crowd environments, and also to get feedback from users, relatives, and nurses.

Classifiers performance on "fall from bed" detection. Times are calculated from the average of four Raspberry Pi. Acc (accuracy), Pre (precision), Rec (recall), F1 (F1 score).

Algorithm	Acc	Pre	Rec	F1	Training time (s)	Classification time (s)
SVM	90.92	0.69	0.81	0.75	26.146	0.472
KNN	91.33	0.69	0.87	0.77	10.340	12.023
GP	87.83	0.62	0.68	0.65	35.237	2.315

Table 3

Matrix confusion for "bed status" experiment.

Ground trut	'n			Evaluati	on		
	On-bed	Sitting	Off-Bed	Ac	Pr	Re	F1
On-bed	120	9	0	0.98	0.93	1.00	0.96
Sitting	0	111	0	0.98	1.00	0.93	0.96
Off-bed	0	0	120	1.00	1.00	1.00	1.00
Avg.				0.98	0.98	0.98	0.97



Fig. 9. HR average and mean absolute error from in-lab experiment.

6.1. In-lab experiment

We carried out the first experiment with 10 participants (6 males and 4 females) from our lab and collaborators. It was done in a controlled environment with the same bed and similar background noise. This experiment was done to tune some algorithm parameters and validate the accuracy of on-bed, off-bed, HR, RR, movement, and posture change. For comparison purposes, we used a commercial fingertip pulse oximeter, Zorvo, to obtain continuous readings of the vital signs (HR and RR). The experiment was set up for 18 min per participant, and it was distributed as follows (i) 3 min on-bed, off-bed, and sitting on the bed (ii) 5 min for HR, (iii) 2 min for RR, (iv) 2 min for movements, and (v) 6 min for posture change.

6.1.1. On-bed, off-bed and sitting experiment

When the person is on-bed, the system starts the algorithms to calculate the vital signs. We collected 36-bed status from each participant, 12 for each type (on-bed, off-bed, sitting). The participants change the status every 5 s according to our directions. In this experiment, the system was set up to detect bed status every 3 s. Table 3 shows the results of the "bed status" test. The method used to calculate the bed status proved to be robust and reliable according to the low error level. The method got a perfect score detecting off-bed because the data-fusion level is high when there is not a periodical signal detected. However, the system detected 9 false on-bed alarms when the participants were seated. Interestingly, after carefully analyzing these specific cases we found that when a person sits at a distance less than 30 centimeters from the sensor, it can detect a faint heartbeat signal that has high auto-correlation levels. We improved the accuracy by analyzing the variability between HR peaks, where high variability implies sitting. In some cases, the system can estimate HR if the person is next to the sensor.

6.1.2. Heart rate experiment

The 5 min HR experiment was divided into two blocks. The first is called "resting mode" where the participant lays on the bed for 3 min in normal condition. The second, "active mode", where the participant needs to do exercise for



Fig. 10. HR visual comparison between Helena and an oximeter. (a) Before exercising. (b) After exercising.



Fig. 11. RR results from controlled respiration using a metronome using a frequency of 20 breath per minute.

2 min before laying on-bed another 2 min to collect high HR and analyze the variation when the person begins to relax. During the experiment, our system was configured to estimate the HR at the same frequency as the oximeter (every 5 s). The error was calculated with 36 readings in rest mode and 24 for the active mode per participant. Fig. 9 shows the HR average and the errors obtained in both modes. The estimated average error was 2.35 bpm (beats per minute) for the resting mode and 2.08 bpm for the active mode. The error results were also affected by the error added for the oximeter that is established in 2 bpm in its description. For example, the maximum error obtained during the experiment was 3.43 bpm, but after calculating it manually using the raw data, the real error was 1.26 bpm.

Fig. 10a shows the estimated HRs from the two devices on resting mode. Here, an error from the oximeter can be appreciated at the time 7:35:00 where there is a drop of 10 bpm, and it is recovered after two readings. Fig. 10b shows the HR behavior during the recovery period of the participant after exercising. Both results show a drop of 20 bpm in the first minute, which is considered normal.

6.1.3. Respiration rate experiment

We used a metronome to validate our respiration rate estimation. The metronome is a generally accepted approach to collect ground truth breathing data in the research community [59,60]. Even though people are forced to breathe artificially using the metronome, we could estimate the accuracy of our method. The metronome was set to sound every three seconds to get a RR of 20 rpm. Forty readings were collected per participant during the two minutes test. Fig. 11 shows the results obtained for each person and the metronome baseline. The average error was 0.52 rpm validating the results obtained in the first experiment. During the HR experiment, the RR was also estimated by both, our system and the oximeter. The purpose was to analyze spontaneous breathing. We obtained 36 and 24 measurement estimations during resting and active mode respectively from each person. The results from the two modes are shown in Fig. 12, having a mean error of 1.56 rpm during the resting mode and 0.61 rpm during the active mode. It is noteworthy that the error in the active mode is smaller because the deep breathing generates a marked difference between the inhalation and exhalation phases improving the envelope calculation.

6.1.4. Movements

The objective of the experiment is body movement detection. However, the experiment was designed to collect data that will be used to develop future features such as the movement location. During it, we asked the participants to



Fig. 12. Results and errors from spontaneous RR calculated by Helena.



Fig. 13. Posture change transition error. B(Back), L(Left), R(Right), C(Chest).

Matrix confusion for posture change detection.

То	Recognized as		Eval
	Change	No change	Acc
Back	110	10	91.67
Chest	115	5	95.83
Right	108	12	90.00
Left	105	15	87.50
Avg.			91.25

move different parts of their bodies one at a time every 5 s. The body part used were the left hand, right hand, left leg, right leg, head, and shoulders. We told the participant which part he/she needed to move according to a pre-established pattern. A total of 240 events were used for calculating false negatives and accuracy. The movements were classified into two groups, "upper movements" (head, arms, and hands) and "lower movements" (legs and feet). The system did not detect 5 movements from the 240 generated, obtaining an accuracy of 97.92%. All of these 5 movements were from lower movements specifically from the feet located outside the bed generating a lower magnitude lower than any on-bed event.

6.1.5. Posture changes

We organized a pattern where the participant had to change to one position from the other three positions. For example, starting from "left", move to "back", then "chest", "back", "right", "back", and so on. We collected 48 posture changes from each participant. The results obtained after the evaluation are presented in Table 4. A total of 438 posture changes were successfully recognized getting an accuracy of 91.25%. Fig. 13 shows the posture change failures from one position to another. As signals from the "back" and "left" have similar characteristics, the posture changes with the highest error were from "back" to the "left" and vice versa. On the other hand, "chest" and "back" posture signals are very different, which leads to fewer errors from "back" to "chest" and vice versa with just one failure.

6.2. Outside experiment

We set up the experiment in 13 different homes with varied environments, beds, and background noise. We got the help from 25 participants. Specifications of the different mattress, size, materials, and distance between floor and

Outside test details and results. Relevant information from the participants, environments, and beds is shown. The results obtained in each of the parts of the system are presented. The highlights are summarized in the last row.

#	House	Gender	Age	Weight	Location	Room location	Floor type under carpet	Bed frame	Mattress material	Mattress Height	Ave-HR \pm error	RR	False On-bed	False Off-bed	Posture change Acc	Fall Acc
1	1	Male	39	196	Bogart	2 Floor	Wood	No	Innerspring	14.5	66bpm \pm 1.32	21.3rpm	0	0	92.86	100
2		Female	3/	154	Bogart	2 Floor	Wood	NO	Memory foam	11.5	$720pm \pm 1.59$	22.2rpm	0	0	90.54	100
3		Male	21	121	Bogart	2 Floor	Wood	No	Memory foam	8	64 bpm ± 2.08	22.2rpm	0	1	-	100
4	2	Male	23	147	Bogart	2 Floor	Wood	No	Memory foam	8	58 bpm ± 2.13	18.1rpm	0	0	-	100
5	2	Female	62	114	Bogart	1 Floor	Concrete	No	Innerspring	10	54bpm \pm 1.27	20.6rpm	0	0	94.44	75
6		Male	19	136	Bogart	1 Floor	Concrete	No	Memory foam	8	76bpm \pm 2.68	16.4rpm	0	0	-	100
7	3	Male	33	172	Athens	4 Floor	Wood	Yes	Hybrid	11	58bpm \pm 1.97	20.1rpm	0	0	87.30	100
8	4	Female	24	127	Athens	2 Floor	Concrete	Yes	Innerspring	9	63 bpm \pm 1.57	19.3rpm	0	0	-	100
9	5	Male	22	143	Athens	2 Floor	Wood	Yes	Innerspring	9	67 bpm \pm 2.73	19.5rpm	0	0	-	100
10		Male	19	204	Alpharetta	2 Floor	Wood	Yes	Hybrid	11	78bpm \pm 1.68	20.0rpm	3	0	-	100
11	6	Male	24	167	Alpharetta	2 Floor	Wood	Yes	Hybrid	11	54 bpm ± 1.58	16.7rpm	1	0	-	100
12		Male	51	154	Alpharetta	2 Floor	Wood	Yes	Hybrid	11	$77bpm \pm 5.28$	20.9rpm	0	0	94.34	100
13		Male	49	200	Alpharetta	2 Floor	Wood	Yes	Memory foam	10	83bpm ± 1.91	16.2rpm	0	0	-	100
14	7	Male	24	183	Alpharetta	1 Floor	Concrete	Yes	Memory foam	10	$63bpm \pm 1.61$	17.4rpm	0	0	92.31	100
15		Female	64	128	Alpharetta	1 Floor	Concrete	Yes	Memory foam	10	$55bpm \pm 1.29$	19.3rpm	0	0	93.33	100
16	8	Male	41	205	Watkinsville	1 Floor	Wood	No	Memory foam	10	69bpm ± 1.95	20.1rpm	0	0	-	100
17	_	Male	25	161	Roswell	3 Floor	Wood	Yes	Hvbrid	14.5	81 bpm ± 5.18	19.7rpm	0	2	-	75
18	9	Female	27	144	Roswell	3 Floor	Wood	Yes	Hybrid	14.5	$74bpm \pm 6.30$	19.1rpm	0	0	-	100
19		Male	33	231	Sandy	2 Floor	Wood	Yes	Memory foam	7	81bpm ± 1.83	20.4rpm	0	0	-	100
20		E	10	140	Springs	0 El		V	M	7	72h	21.2	0	0	01.40	100
20	10	remale	18	145	Salluy	2 FI001	wood	res	welliory toalli	/	720 pm ± 2.18	21.21011	0	0	91.49	100
21	10	Fomalo	50	160	Sandy	2 Floor	Wood	No	Momory form	7	62 hpm ± 150	21 1rpm	0	0		100
21		remate	39	109	Sanuy	2 11001	wood	NU	Welliory Ioani	/	050pm ± 1.59	21.11pm	0	0	-	100
				100	Springs	0.51					241 1 2 2 2	100				
22	11	Male	34	139	Northcross	2 Floor	Wood	Yes	Hybrid	11.5	$64bpm \pm 3.80$	16.2rpm	0	0	-	75
23	12	Male	32	141	Dunwoody	9 Floor	Concrete	Yes	Hybrid	11.5	78bpm \pm 3.24	22.0rpm	1	0	-	100
24	12	Female	44	196	Marietta	2 Floor	Wood	Yes	Innerspring	9.5	75bpm \pm 1.63	18.8rpm	0	0	-	100
25	15	Female	58	146	Marietta	2 Floor	Wood	Yes	Innerspring	9.5	61 bpm ± 1.77	16.7rpm	0	0	-	100
Sumr	nary	Males 16 Females 9	Min 18 Max 64	Min 112 Max 231	Cities 9	Levels 5	Types 2	Yes 17 No 8	Innerspring 6 Memory 11 Hybrid 8	Min 7 Max 14.5	HR Avg. 68.24 Error Avg. 2.41	Min 16.2 Max 22.2	Total 5	Total 4	Avg. 92.08	Avg. 97

HR mean absolute error from Helena and abnormal peaks during estimations.

Average	Average Mean absolute		om
Time	Error±STD	AppleWatch	Helena
1 min	4.28 \pm 2.80 bpm	0.059	0
3 min	4.01 \pm 2.53 bpm	0.183	0
5 min	3.43 \pm 2.23 bpm	0.302	0
10 min	2.89 ± 2.06 bpm	0.611	0
15 min	2.41 \pm 1.88 bpm	0.926	0
30 min	1.76 \pm 1.56 bpm	1.866	0
1 h	1.41 \pm 0.70 bpm	3.384	0

boxes(frames) of the 25 beds are listed in Table 5. The experiment was established for 5 continuous days. Due to some limitation with the number of units and the FDA devices used to validate our results, we experimented during 4 different weeks.

6.2.1. On-bed, off-bed and sitting experiment

We established windows for analyzing on and off-bed. The off-bed window was set between 9 am and 5 pm. Participants confirmed it was their time to be out of their homes. For the on-bed hours, a window was established between 9 pm and 6 am. A limitation of this experiment is that we could not perform sitting validations. However, we analyzed the raw data and the HR and RR results. We did not find any HR and RR calculation outside those expected. We took as expected values for HR between 40 and 140 and respiration between 12 and 25. Also, we considered expected continuous variations between estimations not greater than 8 bpm for HR and 6 for rpm for RR. For example, the HR values of a false on-bed can vary from 62 to 78 bpm in 5 s and for RR from 18 to 26 rpm. We did not find unexpected values during the 125 nights. We confirmed that no HR or RR was estimated while the participants were sitting on the bed. Table 5 shows in two columns the false positives detected per house during the test. The false on-bed column reflects the false positives on-bed or the times when the system detected a person on-bed during off-bed hours. During the entire experiment, five false on-bed detections were presented in two different houses. In house number 6, four false on-bed were detected with a duration smaller than 5 min. We analyzed the signal and realized that the environment was very noisy due to the house construction. Three false on-bed detections occurred in the bed next to the laundry room. During this time we found two mixed signals that correlated as well as HR signal. In another hand, the false on-bed detection on house number 12 was analyzed, and a cardiac signal was found with an HR average of 120 bpm. We found it was from the participant's dog who usually sleeps at the bottom of the bed. The system was able to detect and calculate the dog's HR. The accuracy calculated per day and hour would be 96% and 99.5% respectively. The false off-bed column shows the false exits from the bed. It happens when the system cannot detect the heartbeat due to poor posture and/or a thick object used between the body and the mattress. The accuracy calculated per day and per hour was 97.6% and 99.73%. The solution proposed for both false detection is to recognize a big event before "bed status" change. These events are induced when the person sitting. lying down, leaving the bed, or changing positions. With this approach, the eight detections in total can be reduced to one.

6.2.2. Heart and respiration rate experiment

We used an FDA class 2 medical device, Apple Watch Series 4 [61], to validate the accuracy of our HR method. The Apple Watch provides HR continuous reading in a range between 3 to 10 s. One of the limitations is the battery life since it only allows collecting around 5 h of continuous HR. Participants used an AppleWatch for 5 days at bedtime. We selected it because it is one of the most accurate devices on the market calculating HR with an average error of 2.99 bpm [62]. AppleWatch incorporates some errors during the HR estimation where the changes between 2 consecutive readings are larger than 25 bpm, and it is not normal even is the person is having a panic attack [63]. Fig. 14 shows the graphic results between AppleWatch and our system in different tests. In both figures, our method follows the pattern established by the AppleWatch. Also, the errors from the AppleWatch can be observed. The largest difference during the test is shown at 2:40 am in Fig. 14a. We checked other indicators at the same time, and we observed that in these 12 min, several movements were detected. We can infer that the person was awake at this time. In the same figure, at 3:20 am, there are 10 min of missing results by our system because it detected off-bed status. It was confirmed by the participant.

Table 6 summarizes the results of the entire HR experiment, showing the mean absolute error (MAE) and the peaks introduced by the AppleWatch. Note that the error decreases according to the growth of the time window, and the AppleWatch added more than three errors in one hour. We set 15 min windows for comparison due to many commercial devices calculate HR at this frequency. The MAE of the system is 2.41 bpm even when the AppleWatch incorporates during that time an unreliable peak of data. Devices on the market calculate HR with an average error between 2 and 5 bpm. The error obtained in our experiment shows that our system is within the confidence range estimating HR. Table 7 shows a ranking summary of the MAE using 15 min window. The four best results showed an average error of less than 1.6 bpm. According to Table 5, the common variable to obtain these results is the type of mattress, in this case innerspring. The same variable affects the system performance because in the highest four errors detected, the type of mattresses was



Fig. 14. HR comparison between Apple Watch Version 4 and Helena. (a) Participant 17. (b) Participant 20.

Mean abso	olute error rank of	HR validation.					
Rank	Participant	Mean absolute	Abnormal peaks f	Abnormal peaks from			
		Error±STD	AppleWatch	Helena			
1	C5	1.27 ± 1.29	0.544	0			
2	C15	1.29 ± 1.82	0.742	0			
3	C1	1.32 ± 1.59	1.032	0			
4	C8	1.57 ± 1.26	0.786	0			
22	C22	3.80 ± 3.40	1.143	0			
23	C17	5.18 ± 3.47	0.887	0			
24	C12	5.28 ± 3.90	0.786	0			
25	C18	6.30 ± 4.73	1.446	0			

a hybrid. The hybrid bed has two layers innerspring on the bottom and foam on the top. After checking the signal, we noted that the bed material affects the energy signal. However, there are some exceptions as participants 7, 10 and 11. It is because two other variables influence the energy signal such as person weight and mattress height. In those cases, the participants obtained a lower error than the average due to their weight and heartbeat strength. Even when participant 18 is an overweight person, the error obtained in this experiment was the highest because the participant's bed is thicker.

Validating the respiration rate (RR) was more difficult. Affordable on-the-market devices for monitoring RR have low accuracy and are difficult to wear overnight. For this reason, we only were able to evaluate the results based on the "range of confidence" and the "acceptable variability". The "range of confidence" was established between 13 and 25 rpm as we worked only with healthy participants. The "acceptable variability" means no abrupt changes between one RR estimation and another in a shorter time. None of the participants were outside of the "range of confidence", and the highest variability was obtained by participant 18, with 4 rpm different between two consecutive measurements. Table 5 shows the RR average from each participant.

6.2.3. Posture change and movement experiment

Table 7

We obtained permission from 8 participants to install infrared cameras with motion detection [64]. Fig. 15 shows the camera used and images from the video recorded. The cameras were configured to record video during bedtime between 9 pm to 6 am. Additionally, cam alerts were activated to record 12 s of video after movement detection. The alerts helped us to validate true positives, false positives, and true negatives. In each alert, we verified if there was a posture change or not contrasting it with the results of our system. Additionally, we verified the continuous video recorded in cases where our system registered a posture change, but the alert was not emitted by the camera. This specific case occurred when the person made two continuous posture changes and the alert was activated just the first time.



Fig. 15. Posture change validation. Camera used in the validation and images taken during the experiment. (a) The infrared camera with motion detection. (b) Lady sleeping on the right. (c) Second lady sleeping on the back. (d) Man sleeping on the left.

Table 8 Posture	change validation results.				
Parti	Event detected by the camera	Real posture changes	Changes detected		Acc
			Yes	No	
C1	60	56	52	4	92.86
C2	76	74	67	7	90.54
C5	54	54	51	3	94.44
C7	69	63	55	8	87.30
C12	57	53	50	3	94.34
C14	66	65	60	5	92.31
C15	62	60	56	4	93.33
C22	51	47	43	4	91.49
Avg.					92.08

The test results are shown in Table 8. We noted that the failures produced by the system occurred in 76.31% of cases when the posture change was from left to back or vice versa, similar to results registered in the laboratory test. The average accuracy for the spontaneous posture change is 92.08% which validates results obtained in the simulated posture change test performed in the laboratory. We were unable to perform a quantitative evaluation of movements during the bedtime windows because slight movements detected by the system are not detected by the camera. However, our system detected movements in all alerts not classified as a posture change.

6.2.4. Fall-down

Because falls from bed are not common events to be validated in a real environment, we performed simulated falls when we installed the sensors in different environments. Two falls on each bedside were simulated on each bed. Only 3 events were not classified as a fall from bed having an accuracy of 97%. Throughout the experiment, we did not detect any false fall from the bed. Even when the method was trained with different fall positions, people, environments, and beds, the data were simulated falls. For this reason, we cannot guarantee the same level of accuracy in spontaneous falls. However, we plan to integrate another sensor installed on the floor to combine the events that occur in the bed and on the floor, to provide a higher level of confidence [13].

6.3. Senior care center test

Because two previous tests are performed with healthy people and under 60 years, we decided to perform an additional test with senior people. Two units were installed for two days in the Magnolia Senior Care facility located in the city of Loganville GA. We received approval from the manager, the participants, and their families. The participants were 87 years old and 147 pounds gentleman, and a 74 years old and 132 pounds lady. The on-bed, off-bed and sitting status, were confirmed by the nurses. The HB and RR were monitored in real-time from our central dashboard. The values of both HR and RR were normal throughout the process. No false fall from bed was detected. We could not perform posture change or movement validations. Fig. 16 shows the monitoring dashboard during part of the process, which allows seeing the report in real-time showing the last 5 min results and also to check historical results.

7. Discussion

In this section, we explore the drawbacks of the proposed system and the potential solutions to those drawbacks. We plan to follow the presented recommendation to improve the system and potentially make it commercially available.





7.1. Drawbacks

Several aspects need to be address to ensure a reliable and accurate system as Helena, especially aspects related to environment interference.

7.1.1. Location of the sensor

In the Helena system, the sensor must be placed within 40 centimeters from the heart location when the person lays down on the bed. This is a requirement mostly because the signal degrades due to multiple materials on the bed. We conducted tests on king, queen, and single beds, and as long as the sensor is placed correctly (<40 cm of the heart location), the application works as expected. However, in a king bed, if the sensor is placed on one side and the person sleeps on the other side, the signal is not going to provide the expected results, the application triggers multiple false alarms, and the measurement of heart and respiration are not correct. This is similar to other commercial sleep pads, such as the Nokia Sleep Pad [7], which has to be placed under the back of the person.

7.1.2. Number of people in bed at the same time

The design of the Helena system was conceived to work on a bed with only one occupant. Currently, the system does not manage the separation of the signals from multiple people. To address this issue, the authors need to add more than one sensor to the system and apply a blind separation technique to differentiate the vibration signal from different people. This study has been reserved for future work.

7.1.3. Sensitivity to external artifacts

During the initial tests on uncontrolled environments, we realized the multiple external artifacts that can interfere with the vibration signal. One of the most common ones is the Air Conditioner (AC) vibration. To solve this problem, we applied a narrow-band notch filter as explained in Section 5.1. The filter works as expected because the vibration of the AC is constant in frequency. However, there are other artifacts like vibrations from washers and dryers that are more challenged due to the variant nature of the washer and dryer cycles. Those interferences were also reported in the experiment section. For future work, we plan to conduct a deep study to design an advanced adapted filter that allows us to mitigate the effect of these variant artifacts. Furthermore, we are going to study the possibility to incorporate a light learning process for filter adaptation.

7.2. Future improvements

In Section 5.3, we explain how we conduct the movement and posture change detection. The posture change was referred to as the change in the position of the body, but not consider specifically which position is the detected one. Understanding which the current posture of the person in bed could potentially provide information to study medical concerns such as quality of sleep, apnea, breathing disorders, sleep paralysis, gastroesophageal reflux, body sores, etc. [65–71]. To improve this detection, added some machine learning techniques to have a more robust and more accurate posture change algorithm. Information from the stages and the complete heartbeat motion cycles (events) are used to get the features used for our posture identification model. The heartbeat motion features have been extracted for user authorization [58]. Similarly, in our study, the bed vibrations caused by heartbeat and respiration captured by a bedmounted seismometer can be used to get physical human vibration features to differentiate the sleep postures on the bed. A heartbeat motion cycle has seven stages, atrial contraction (ATC); mitral valve closing (MC); aortic valve opening (AO); point of maximal acceleration in the aorta (MA); aortic valve closure (AC); mitral valve opening (MO); rapid filling of left ventricle (RF), that are shown on Fig. 17.



Fig. 17. The heartbeat related bed vibrations are produced by the consecutive mechanical movements during a heartbeat (MC, IM, AO, then MA). The relative amplitudes change depending on the person and the sleep postures. So the movement features can be used for posture detection.



Fig. 18. Sleep posture identification. (a) Extracted template. (b) vibration data. (c) stage isolation based on cross-correlation coefficients. (d) zoom-in display.

We observe that different postures on the bed cause different variation patterns in the time domain amplitude of some motion stages in Fig. 4. Information from the stages and the complete heartbeat motion cycles (events) are used to get the features used for our posture identification model. The two steps of our feature extraction algorithm shown in Fig. 18 are described as follows:

7.2.1. Segmentation

We separate the constant signal read by the seismometer into individual separated events. We implement the crosscorrelation between the bed seismic signal and a template of the reference heartbeat motion cycle. The correlation coefficient is calculated, and a threshold of 0.90 is established to identify the starting point of the events. The result of the correlation is shown in Fig. 18(c) Then, the event is segmented based on the starting point and the template length.

7.2.2. Features selection

Once the event is isolated, we locate the local maximum points by their amplitudes (peaks). To ignore some peaks presented by the noise, we set a minimum distance between peaks. We realize by the observation that some stages (AO, AL, and RF) change their magnitude on time domain according to the position on the bed. We use these changes to construct three features for the classification model. The first and the second features are rate magnitude on time-domain between AO:RF and AL:RF. Also, we realize that there are some delays in the arrival time of RF depending on the lying position. Therefore, we use as the third feature the time difference between ATC (the first stage) and RF. Finally, we incorporate the event standard deviation as a feature to increase the accuracy of our classification model.

7.2.3. Sleep posture identification

To identify the sleep postures, we propose to use a machine learning-based approach. Based on a supervised machine learning algorithm, Support Vector Machine (SVM) [57], data are transformed using a kernel to find the optimal boundary between the different classes. SVM and its derivatives have been widely used in the biomedical engineering [13,58] because it is efficient and easy to implement on embedded devices.

Further experiments on this new posture change methodology are going to be considered in future work and incorporate into the Helena system for a new systematic experimental round with multiple people during the whole night of sleep.

8. Conclusion

An accurate and suitable pervasive sleep monitoring system (Helena) based on a contactless and non-invasive bedmounter sensor is proposed in this paper. Novel methods for heart rate and respiration rate estimations have been introduced only based on heart vibrations on the bed. For the first time, a non-intrusive method to detect and alert falls from bed has been introduced with an accuracy higher than 95%. The system also allows real-time monitoring of movements and posture changes. The system is able to estimate heart and respiration rate every 3 s, and changes in posture, movements, and fall downs with a delay of a maximum of 5 s. The alerts are then sent to caregivers, which enable early intervention and rapid decision-making. While the real-time feature may not be urgent for movement and posture change detection, this feature is significant in terms of fall detection, as this type of event requires immediate attention. The performance has been tested multiple times on different people, beds, environments to determine its accuracy and has been compared with FDA approved wearable devices reaching promising results. The system is envisioned to be a life-changing for high-risk communities, like seniors, where there is a lack of non-contact devices for monitoring their sleep patterns and assess their vital signs.

CRediT authorship contribution statement

Maria Valero: Conceptualization, Methodology, Algorithm design, Software development, Document preparation. Jose Clemente: Conceptualization, Methodology, Algorithm design, Implementation. Fangyu Li: Conceptualization, Signal study, Algorithm design, Noise removal, Experiment. WenZhan Song: Conceptualization, Guidance.

Declaration of competing interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

References

- The World Bank Data, World bank's total population and age/sex distributions of the united nations population, 2017, URL https://data. worldbank.org/indicator/SP.POP.65UP.TO.
- [2] A. Roberts, S. Ogunwole, L. Blakeslee, M. Rabe, The population 65 years and older in the United States (ACS-38), 2016.
- [3] R.B. Berry, R. Brooks, C.E. Gamaldo, S.M. Harding, C.L. Marcus, B.V. Vaughn, et al., The aasm manual for the scoring of sleep and associated events, in: Rules, Terminology and Technical Specifications, 176, American Academy of Sleep Medicine, Darien, Illinois, 2012.
- [4] C.A. Kushida, A. Chang, C. Gadkary, C. Guilleminault, O. Carrillo, W.C. Dement, Comparison of actigraphic, polysomnographic, and subjective assessment of sleep parameters in sleep-disordered patients, Sleep Med. 2 (5) (2001) 389–396.
- [5] G. Nilius, U. Domanski, M. Schroeder, K.-J. Franke, A. Hogrebe, L. Margarit, M. Stoica, M.-P. d'Ortho, A randomized controlled trial to validate the alice pdx ambulatory device, Nat. Sci. Sleep 9 (2017) 171.
- [6] S. Grover, I. Bajwa, A. Butchko, et al., A comparison of traditional laboratory-based polysomnography and the alice pdx portable monitoring device, and the usability of the alice pdx, 2011, available at: http://www.healthcare.philips.com/asset.aspx (Accessed February 2, 2011).
- [7] Nokia, Nokia withings sleep pad, 2019, https://www.withings.com/us/en/sleep (Last Accessed: 2019-04-30).
- [8] EarlySense, Early sense, 2019, https://www.earlysense.com/ (Last Accessed: 2019-04-1).
- [9] M. Tenhunen, E. Elomaa, H. Sistonen, E. Rauhala, S.-L. Himanen, Emfit movement sensor in evaluating nocturnal breathing, Respir. Physiol. Neurobiol. 187 (2) (2013) 183–189.
- [10] F. Li, J. Clemente, M. Valero, Z. Tse, S. Li, W. Song, Smart home monitoring system via footstep-induced vibrations, IEEE Syst. J. 14 (3) (2020) 3383-3389.
- [11] J. Clemente, W. Song, M. Valero, F. Li, X. Li, Indoor person identification and fall detection through non-intrusive floor seismic sensing, in: 5th IEEE International Conference on Smart Computing(SMARTCOMP), 2019, pp. 417–424.
- [12] S. Pan, T. Yu, M. Mirshekari, J. Fagert, A. Bonde, O.J. Mengshoel, H.Y. Noh, P. Zhang, Footprintid: Indoor pedestrian identification through ambient structural vibration sensing, Proc. ACM Interact. Mob. Wearable Ubiquitous Technol. 1 (3) (2017) 89.
- [13] J. Clemente, F. Li, M. Valero, W. Song, Smart seismic sensing for indoor fall detection, location and notification, IEEE J. Biomed. Health Inf. (2019) Online.
- [14] R.N. Aurora, S.P. Patil, N.M. Punjabi, Portable sleep monitoring for diagnosing sleep apnea in hospitalized patients with heart failure, Chest (2018).
- [15] F. Li, J. Clemente, W. Song, Non-intrusive and non-contact sleep monitoring with seismometer, in: 2018 IEEE Global Conference on Signal and Information Processing (GlobalSIP), IEEE, 2018, pp. 449–453.
- [16] P. Kowallik, W. Hubert, Method and device for sleep monitoring, 2004, US Patent 6, 752, 766.
- [17] K.-L. Hsiao, What drives smartwatch adoption intention? Comparing apple and non-apple watches, Libr. Hi Tech. 35 (1) (2017) 186-206.
- [18] L. Jeon, J. Finkelstein, Consumer sleep tracking devices: a critical review, in: Digital Healthcare Empowering Europeans: Proceedings of MIE2015, Vol. 210, 2015, p. 458.
- [19] E. Jovanov, Preliminary analysis of the use of smartwatches for longitudinal health monitoring, in: 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), IEEE, 2015, pp. 865–868.
- [20] K.M. Diaz, D.J. Krupka, M.J. Chang, J. Peacock, Y. Ma, J. Goldsmith, J.E. Schwartz, K.W. Davidson, Fitbit[®]: An accurate and reliable device for wireless physical activity tracking, Int. J. Cardiol. 185 (2015) 138–140.
- [21] H. Maki, H. Ogawa, S. Tsukamoto, Y. Yonezawa, W.M. Caldwell, A system for monitoring cardiac vibration, respiration, and body movement in bed using an infrared, in: Engineering in Medicine and Biology Society (EMBC), 2010 Annual International Conference of the IEEE, IEEE, 2010, pp. 5197–5200.
- [22] M. Alaziz, Z. Jia, J. Liu, R. Howard, Y. Chen, Y. Zhang, Motion scale: A body motion monitoring system using bed-mounted wireless load cells, in: Connected Health: Applications, Systems and Engineering Technologies (CHASE), 2016 IEEE First International Conference on, IEEE, 2016, pp. 183–192.
- [23] A.M. Adami, T.L. Hayes, M. Pavel, C.M. Singer, Detection and classification of movements in bed using load cells, in: Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the, IEEE, 2006, pp. 589–592.

- [24] C. Brüser, A. Kerekes, S. Winter, S. Leonhardt, Multi-channel optical sensor-array for measuring ballistocardiograms and respiratory activity in bed, in: Engineering in Medicine and Biology Society (EMBC), 2012 Annual International Conference of the IEEE, IEEE, 2012, pp. 5042–5045.
- [25] H. Ni, B. Abdulrazak, D. Zhang, S. Wu, Unobtrusive sleep posture detection for elder-care in smart home, in: International Conference on Smart Homes and Health Telematics, Springer, 2010, pp. 67–75.
- [26] S. Pan, N. Wang, Y. Qian, I. Velibeyoglu, H.Y. Noh, P. Zhang, Indoor person identification through footstep induced structural vibration, in: Proceedings of the 16th International Workshop on Mobile Computing Systems and Applications, ACM, 2015, pp. 81–86.
- [27] Z. Jia, M. Alaziz, X. Chi, R.E. Howard, Y. Zhang, P. Zhang, W. Trappe, A. Sivasubramaniam, N. An, Hb-phone: a bed-mounted geophone-based heartbeat monitoring system, in: Information Processing in Sensor Networks (IPSN), 2016 15th ACM/IEEE International Conference on, IEEE, 2016, pp. 1–12.
- [28] Z. Jia, A. Bonde, S. Li, C. Xu, J. Wang, Y. Zhang, R.E. Howard, P. Zhang, Monitoring a person's heart rate and respiratory rate on a shared bed using geophones, 2017.
- [29] Geophone SM-24, 2020, https://www.sparkfun.com/products/11744 (Last Access: 10/30/2020).
- [30] Ti lmv358, 2020, https://www.ti.com/product/LMV358 (Last Access: 10/30/2020).
- [31] G.E. Box, G.M. Jenkins, G.C. Reinsel, G.M. Ljung, Time Series Analysis: Forecasting and Control, John Wiley & Sons, 2015.
- [32] T. O'Haver, Peak finding and measurement, 2009.
- [33] J. Park, H. Cho, R.K. Balan, J. Ko, Heartquake: Accurate low-cost non-invasive ecg monitoring using bed-mounted geophones, Proc. ACM Interact. Mob. Wearable Ubiquitous Technol. 4 (3) (2020) 1–28.
- [34] Y. LeCun, Y. Bengio, G. Hinton, Deep learning, Nature 521 (7553) (2015) 436-444.
- [35] T. Choudhary, M. Bhuyan, K. Bora, L. Sharma, Design of breathing-states detector for m-health platform using seismocardiographic signal, 2020, arxiv preprint arxiv:2002.10510.
- [36] T. Choudhary, M. Bhuyan, L. Sharma, Effect of respiratory effort levels on scg signals, in: 2019 IEEE Region 10 Symposium (TENSYMP), IEEE, Kolkata, India, 2019, pp. 143–146.
- [37] T. Choudhary, L. Sharma, M.K. Bhuyan, Automatic detection of aortic valve opening using seismocardiography in healthy individuals, IEEE J. Biomed. Health Inf. 23 (3) (2018) 1032–1040.
- [38] L. Leppäkorpi, Beddit is a new kind of device and app for tracking & improving sleep and wellness, 2014, Beddit Is a New Kind of Device and App for Tracking & Improving Sleep and Wellness.
- [39] M. Helfand, V. Christensen, J. Anderson, Technology Assessment: Early Sense for Monitoring Vital Signs in Hospitalized Patients, Department of Veterans Affairs (US), 2016.
- [40] S. Blake, S. Pittman, M. MacDonald, K. Sun, B. Lanzi, D. Clark, L. Hueser, S. Fabregas, J. Shambroom, D. White, Assessment of a wireless dry sensor to detect sleep in healthy volunteers and subjects with sleep disorders, in: Sleep, 32, AMER ACAD SLEEP MEDICINE ONE WESTBROOK CORPORATE CTR, STE 920, WESTCHESTER ..., 2009, p. A370.
- [41] T. Althoff, E. Horvitz, R.W. White, J. Zeitzer, Harnessing the web for population-scale physiological sensing: A case study of sleep and performance, in: Proceedings of the 26th International Conference on World Wide Web, International World Wide Web Conferences Steering Committee, 2017, pp. 113–122.
- [42] P. USA, Philips alice pdx, 2019, https://www.usa.philips.com/healthcare/product/HC1043941/alice-pdx-portable-sleep-diagnostic-system (Last Accessed: 2019-04-30).
- [43] WatchPad, Itamar medical watchpat simple sleep health management, 2019, https://www.itamar-medical.com/watchpat-main/ (Last Accessed: 2019-04-18).
- [44] K. Nakajima, Y. Matsumoto, T. Tamura, Development of real-time image sequence analysis for evaluating posture change and respiratory rate of a subject in bed, Physiol. Meas. 22 (3) (2001) N21.
- [45] S. Okada, Y. Ohno, K. Kato-Nishimura, I. Mohri, M. Taniike, et al., Examination of non-restrictive and non-invasive sleep evaluation technique for children using difference images, in: 2008 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, IEEE, 2008, pp. 3483–3487.
- [46] SleepCycle USA, Sleep cycle alarm clock, 2019, https://www.sleepcycle.com/ (Last Accessed: 2019-04-30).
- [47] ResMed, Resmed s+ sleep monitoring, 2019, https://www.resmed.com/us/en/consumer/s-plus.html (Last Accessed: 2019-04-18).
- [48] Sleeppace, Sleeppace smart bedding, 2019, http://www.sleepace.com/smartBedding.html (Last Accessed: 2019-04-18).
- [49] Smart Bedding, Sleepace sleep tracking pad, 2019, http://www.sleepace.com/tracking_pad.html (Last Accessed: 2019-04-18).
- [50] Early Sense Live, Live by early sense, 2019, https://techcarehq.com/posts/earlysense-live (Last Accessed: 2019-04-1).
- [51] C.E. Krohn, Geophone ground coupling, Geophysics 49 (6) (1984) 722-731.
- [52] Weihai Sunfull Geophysical Exploration Equipment Co., Ltd, Geophone, 3-Component, http://www.sunfull.com/content/?148.html.
- [53] E. Upton, Raspberry pi 3, 2016, URL https://www.raspberrypi.org/products/raspberry-pi-3-model-b.
- [54] Influxdata Inc, Influxdb, 2019, URL https://www.influxdata.com/.
- [55] Grafana Labs, Grafana, 2018, URL https://grafana.com/.
- [56] R.S. Holambe, M.S. Deshpande, Nonlinear measurement and modeling using teager energy operator, in: Advances in Non-Linear Modeling for Speech Processing, Springer, 2012, pp. 45–59.
- [57] C. Cortes, V. Vapnik, Support-vector networks, Mach. Learn. 20 (3) (1995) 273-297.
- [58] L. Wang, K. Huang, K. Sun, W. Wang, C. Tian, L. Xie, Q. Gu, Unlock with your heart: Heartbeat-based authentication on commercial mobile phones, Proc. ACM Interact. Mob. Wearable Ubiquitous Technol. 2 (3) (2018) 140.
- [59] A. Johansson, Neural network for photoplethysmographic respiratory rate monitoring, Med. Biol. Eng. Comput. 41 (3) (2003) 242–248.
- [60] R. Ravichandran, E. Saba, K.-Y. Chen, M. Goel, S. Gupta, S.N. Patel, Wibreathe: Estimating respiration rate using wireless signals in natural settings in the home, in: 2015 IEEE International Conference on Pervasive Computing and Communications (PerCom), IEEE, 2015, pp. 131–139.
 [61] Apple, Applewatch4, 2019, URL https://www.apple.com/apple-watch-series-4/ (Last Accessed: 2019-04-18).
- [62] E.A. Thomson, K. Nuss, A. Comstock, S. Reinwald, S. Blake, R.E. Pimentel, B.L. Tracy, K. Li, Heart rate measures from the apple watch, fitbit charge hr 2, and electrocardiogram across different exercise intensities, J. Sports Sci. 37 (12) (2019) 1411–1419.
- [63] K. Spiegelhalder, M. Hornyak, S.D. Kyle, D. Paul, J. Blechert, E. Seifritz, J. Hennig, L. Tebartz van Elst, D. Riemann, B. Feige, Cerebral correlates of heart rate variations during a spontaneous panic attack in the fmri scanner, Neurocase 15 (6) (2009) 527–534.
- [64] Wyze, Wyze2, 2019, URL https://www.wyze.com/product/wyze-cam-v2/ (Last Accessed: 2019-04-18).
- [65] V.K. Kapur, D.H. Auckley, S. Chowdhuri, D.C. Kuhlmann, R. Mehra, K. Ramar, C.G. Harrod, Clinical practice guideline for diagnostic testing for adult obstructive sleep apnea: an American academy of sleep medicine clinical practice guideline, J. Clin. Sleep Med. 13 (3) (2017) 479–504.
- [66] R.-S. Hsiao, Z. Mi, B.-R. Yang, L.-J. Kau, M.A. Bitew, T.-Y. Li, Body posture recognition and turning recording system for the care of bed bound patients, Technol. Health Care 24 (s1) (2016) S307–S312.

M. Valero, J. Clemente, F. Li et al.

- [67] J.B. Lee, Y.H. Park, J.H. Hong, S.H. Lee, K.H. Jung, J.H. Kim, H. Yi, C. Shin, Determining optimal sleep position in patients with positional sleep-disordered breathing using response surface analysis, J. Sleep Res. 18 (1) (2009) 26–35.
- [68] J.A. Cheyne, Situational factors affecting sleep paralysis and associated hallucinations: position and timing effects, J. Sleep Res. 11 (2) (2002) 169–177.
- [69] M. Van Herwaarden, D.A. Katzka, A. Smout, M. Samsom, M. Gideon, D. Castell, Effect of different recumbent positions on postprandial gastroesophageal reflux in normal subjects, Amer. J. Gastroenterol. 95 (10) (2000) 2731–2736.
- [70] J. De Koninck, P. Gagnon, S. Lallier, Sleep positions in the young adult and their relationship with the subjective quality of sleep, Sleep 6 (1) (1983) 52–59.
- [71] P. Martin-Du, R. Benoit, L. Girardier, The role of body position and gravity in the symptoms and treatment of various medical diseases, Swiss Med. Wkly. 134 (3738) (2004).



Maria Valero is assistant professor in the College of Computing and Software Engineering, Department of Information Technology at Kennesaw State University. She received her Ph.D in Electrical and Computer Engineering at University of Georgia in 2019 and Master in Computer Science from University of Los Andes (Venezuela) in 2009. She is professor at University of Tachira (Venezuela) since 2004. Her research interests include distributed computing, signal processing, wireless sensor networks, Internet of Things (IoT), and cyber–physical systems.



Jose Clemente received his Ph.D. in Electrical and Computer Engineering at University of Georgia in 2019. He received his MS in Computer Science in 2009 at the University of Los Andes (Venezuela) and his bachelor in Computer Engineering from the University of Tachira (Venezuela). Currently, he is a research scientist and software engineer at The Home Depot. His research interests are fog and edge computing, cloud computing, quality of service, sensor networks and cyber–physical systems.



Fangyu Li received the B.S. and M.S. degrees in electrical engineering from Beihang University, Beijing, China, and Tsinghua University, Beijing, China, in 2009 and 2013, respectively, and the Ph.D. degree in geophysics from The University of Oklahoma, Norman, OK, USA, in 2017. Between 2017 and 2020, Dr. Li did his postdoctoral research at the School of Electrical and Computer Engineering at University of Georgia. He is currently an assistant professor at the Department of Electrical and Computer Engineering at Kennesaw State University. His current research interests include signal processing, seismic imaging, machine learning, deep learning, distributed computing, smart healthcare, Internet of Things (IoT), and cyber-physical systems (CPS).



WenZhan Song received Ph.D. in Computer Science from Illinois Institute of Technology (2005), B.S. and M.S. degree from Nanjing University of Science and Technology (1997 and 1999). He is a Chair Professor of Electrical and Computer Engineering in the University of Georgia. Dr. Song's research focuses on cyber–physical systems and security and their applications in energy, environment, food and health sectors. He received National Science Foundation (NSF) CAREER award in 2010.